

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

PORTLAND STONE PORTLAND
LOCATION

Machine Details

Cust ref	
Make	BILANCIAI
Model	DD700
Ser No	24361008
Capacity	50000 / kg /
e/d	20 / ks /
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
PDG 1-22	T212668	31/12/24

Cert Interval	12 MONTHS
Next Cert	FEB 2026

Tolerance +/-	kg
0- 12000	20
10000-40000	40
40000-50000	60

Comments	
NEW VERIFICATION	
FILIFT = 5540	
UNIT = 19.160	

2	4	6
1	3	5
x		

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0 kg	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
25540	25540	0	25540	0
34700	34700	0	34700	0
40700	40700	0	40700	0
44700	44700	0		

Repeatability

Load	Ind	Ind	Ind

Repeatability

Load	Ind	Ind	Ind
40700	40700	40700	40700

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Eccentric Load 10000

Reading	Dev	Reading	Dev
1	10000	2	10000
3	10000	4	10000
5	10000	6	10000
7		8	

Signature  Name S. THOMPSON Date 10/12/25
C1.3.88.1N

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

PORTLAND STONE PORTLAND LOCATION
--

Machine Details

Cust ref	
Make	BILANCIAI
Model	DD700
Ser No	24180565
Capacity	50000 / KG /
e/d	20 / KG /
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
PQG 1-22	TR12665	3/12/24

Cert Interval	12 MONTHS
Next Cert	FEB 2026

Tolerance +/-	kg
0- 10000	20
10000-40000	40
40000-50000	60

Comments NEW VERIFICATION FLIFT = 5540 UNIT = 19160	<table border="1" style="margin: auto;"> <tr> <td>2</td> <td>4</td> <td>6</td> </tr> <tr> <td>1</td> <td>3</td> <td>5</td> </tr> <tr> <td colspan="3" style="text-align: right;">X</td> </tr> </table>	2	4	6	1	3	5	X		
2	4	6								
1	3	5								
X										

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0kg	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
25540	25540	0	25540	0
34700	34700	0	34700	0
40700	40700	0	40700	0
44700	44680	-20		

Repeatability

Load	Ind	Ind	Ind

Repeatability

Load	Ind	Ind	Ind
40700	40700	40700	40700

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Eccentric Load 10000 kg

Reading	Dev	Reading	Dev
1	10000	2	10000
3	10000	4	10000
5	10000	6	10000
7		8	

Signature Name S. THOMPSON Date 10/2/25