

Weightron Bilanciai Ltd,
 Titan Works
 Bridge Way off Broombank Road,
 Chesterfield Trading Estate,
 Chesterfield S41 9QJ, UK
 Tel: 01246 260062
 Fax: 01246 260844
 www.weightron.com

Weightron Bilanciai Ltd,
 Unit 4b, Block 4
 Westpark, Chelston
 Wellington TA21 9AD, UK
 Tel: 01823 662355
 Fax: 01823 6621017
 www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the Calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

PORTLAND STONE - COOMBEFIELD QUARRY	
Portland Dorset DT5 2DH	
Location	Site

Machine Details

Cust ref	Weighbridge	
Make	Bilanciai	
Model	DD700 ABS Digital Indicator	
Serial No	T24361008	
Capacity	50000	kg
e/d	20	kg
Working cap	44000	

Test Weights

Set No.	Cert No.	Date Cal
PDG	TR13167	19/02/2027

Cert Interval	12 Months
Next Cert	Feb 2027

Tolerance +/-	kg
0 - 10000	20
10000 - 40000	40
40000 - 50000	60

Comments

AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
2000	2000	0	2000	0
10000	10000	0	10000	0
16000	16000	0	16000	0
20000	20000	0	20000	0
24500	24500	0	24500	0
40500	40500	0	40500	0
44500	44500	0	44500	0

DEFINITIVE LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev

Repeatability

Load kg	Ind	Ind	Ind
40500	40500	40500	40500

Repeatability

Load kg	Ind	Ind	Ind

Eccentric Load 10000

Reading	kg	Dev	Reading	kg	Dev
1	10000	0	2	10000	0
3	10000	0	4	10000	0
5			6		
7			8		
9			10		
11			12		

Eccentric Load

Reading	kg	Dev	Reading	kg	Dev
1			2		
3			4		
5			6		
7			8		
9			10		
11			12		

Signature:

Name: Harry Gregory

Date: 19/02/2026 09:58

Weightron Bilancai Ltd,
 Titan Works
 Bridge Way off Broombank Road,
 Chesterfield Trading Estate,
 Chesterfield S41 9QJ, UK
 Tel: 01246 260062
 Fax: 01246 260844
 www.weightron.com

Weightron Bilancai Ltd,
 Unit 4b, Block 4
 Westpark, Chelston
 Wellington TA21 9AD, UK
 Tel: 01823 662355
 Fax: 01823 6621017
 www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the Calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

PORTLAND STONE - COOMBEFIELD QUARRY	
Portland Dorset DT5 2DH	
Location	Site

Machine Details

Cust ref	Weighbridge	
Make	Bilanciai	
Model	DD700 ABS Digital Indicator	
Serial No	T24180868	
Capacity	50000	kg
e/d	20	kg
Working cap	44000	

Test Weights

Set No.	Cert No.	Date Cal
PDG	TR13167	19/02/2027

Cert Interval	12 Months
Next Cert	Feb 2027

Tolerance +/-	kg
0 - 10000	20
10000 - 40000	40
40000 - 50000	60

Comments

AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
2000	2000	0	2000	0
10000	10000	0	10000	0
16000	16000	0	16000	0
20000	20000	0	20000	0
24500	24500	0	24500	0
40500	40500	0	40500	0
44500	44500	0	44500	0

DEFINITIVE LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev

Repeatability

Load kg	Ind	Ind	Ind
40500	40500	40500	40500

Repeatability

Load kg	Ind	Ind	Ind

Eccentric Load 10000

Reading	kg	Dev	Reading	kg	Dev
1	10000	0	2	10000	0
3	10000	0	4	10000	0
5			6		
7			8		
9			10		
11			12		

Eccentric Load

Reading	kg	Dev	Reading	kg	Dev
1			2		
3			4		
5			6		
7			8		
9			10		
11			12		

Signature:

Name: Harry Gregory

Date: 19/02/2026 11:02